



<https://www.abrj.org/>

Impact Factor 2.30

5G-Enabled Remote Neuroscience: Low-Latency Neural Data Transmission for Tele-Neurorehabilitation

Author Details: Sayma Nasrin Shompa

International Islamic University Chittagong.

Computer Science and Engineering, Bangladesh, Chittagong,

Email: Saimanasrin453@gmail.com

<https://orcid.org/0009-0004-8637-1108>

Abstract:

Integration of the fifth generation (5G) wireless technology in neurosciences provides a transformative solution for the boundaries of the current tele-neurorehabilitation system, which rely on 3G/4G or Wi-Fi and suffer from high detailed, low bandwidth and incredible data transmission. This paper proposes 5G-enabled tele-neurorehabilitation framework-5G-Saksham-which ultra-widely low-stamped communication (URLLC), Multi-Access Edge Computing (MEC), and Network Slicing Close Lupp Takes advantage for. High-dimensional brain-computer interfaces (BCIs) and EEG data are streamflow with minimal delays, processed locally on the edge, and used to give adaptive response through the Emarsiv AR/VR environment. Performance analysis and simulated pilot study reduction in delay (<10 mS), signal fidelity (> 95%), and patient engagement show significant improvement in engagement, leading to motor recovery results and medical compliance compared to traditional approaches. Additionally, the framework addresses scalability, security, and interoperability challenges, which paves the way for patients in remote or undescribed areas, accessible to neurorehabilitation of clinic-quality, accessible to patients. Our findings unlock the position, continuous, data-driven and patient-focused neurological care as an important environ for the next generation of 5G tele-neurosciences

Keywords: 5G, neuroscience, tele-neurorehabilitation, brain-computer interface, edge computing, low-latency

I. Introduction

Integration of 5G technology in neuroscience promises a transformative leap in tele-neuroeducation, where patients with neurological losses receive remote therapy supported by real-time nerve data transmission. Traditional telemedicine encounters delay and bandwidth obstacles that limit the efficacy of transmitting complex brain signals [2]. 5g, with its ultra-low delay and high bandwidth, provides a solution that enables spontaneous, real-time remote neuroscience applications [3,12]. This advancement can revolutionize neuroeducation by making remedies more accessible and responsible. Tele-neurorehabilitation provides a promising option for traditional clinic-based care, yet its efficacy has been forced by the boundaries of the traditional wireless network [28]. The real-time brain-computer interface (BCI) applications for the high-dimensional, time-sensitive nature ultra-low delay and high bandwidth for applications for applications demands 3G/4G and Wi-Fi infrastructure in the current strength [4,21]. This paper examines the conversion capacity of the fifth generation (5G) wireless technology to overcome these obstacles. We propose a framework for spontaneous remote neurosciences by taking advantage of the 5G ultra-wisdom low-latency communication (URLLC), multi-gigabit throughput, and edge computing [12,35,36]. Our analysis indicates that 5G strong, real-time closed-loop enables neurophydback, emergent AR/VR facilitates integration, and significantly improves the access and quality of distance rehabilitation [27,31]. We conclude that 5G is an

important enabler for the next generation of tele-neurorehabilitation, capable of supporting scalable, adaptive remedies, which promises the patient's consequences and maximum equity in neurological care.

Neurological disorders such as strokes, painful brain injuries, and neurocognitive diseases represent an important global health burden, often releasing patients with prolonged motor, cognitive and sensory loss [19]. Traditional neurorehabilitation therapy requires in-practice sessions with specialized equipment and clinical supervision, which can limit access to patients especially in remote or underserved areas. Tele-neuroeducation, which enables distance distribution of therapy through digital platforms, has emerged as a promising solution to solve these challenges by expanding rehabilitation services beyond clinical settings [28]. However, current telemedicine infrastructure facing 3G, 4G, or Wi-Fi networks boundaries that obstruct their effectiveness in neurological applications [21]. Nerve signals such as electroencephalography (EEG) or brain-computer interfaces (BCI) data are high-dimensional and low-density, high-bandwidth transmission is required to support real-time analysis and closed-loop response [4]. The delay or data loss during transmission can compromise the quality of nervous information, reduce medical efficacy and patient engagement [21].

The arrival of the fifth generation (5G) wireless network provides transformational capacity for distance neurological [3,12]. With Ultra-Low latency (low as 1 MS), enhances the increase in reliability, and significantly higher throughputs than the previous generations, capable of uninterrupted transmission of 5G nerve data currents [35]. This capacity is especially important for tele-neurorehabilitation, where real-time accountability allows adaptive remedies such as brain stimulation or immersive virtual reality (VR)-in-place rehabilitation exercises [27]. In addition, characteristics such as edge computing and network slicing strengthen 5G suitability by localizing data processing, reducing network congestion and ensuring secure distribution of neurorehabilitation services [29,36].

This paper examines the role of 5G in enabling distance neurological applications with focus on tele-narratability. We discuss the boundaries of existing network technologies, review the related functions in brain-computer interfaces and network neurorehabilitation, and introduce a 5G-able system framework that integrates nerve data acquisition, transmission and aging. The proposed structure is evaluated through performance comparison with the heritage network, highlighting its ability to improve accountability, medical results and access to access. Finally, we solve the challenges of data protection, device inequality, and protocol standardization, underlining future directions for 5G-powered neurological science. Remote neurosciences involve capturing and transmitting neural data-like EEG, intracranial recording, or brain-computer interface (BCI) for signal-diagnosis, surveillance and therapy. These data are often high dimensional and require rapid processing to be clinically useful. Traditional wireless network (3G/4G/Wi-Fi) introduces delays and packet loss that can degrade the quality of the transmitted nerve signals, possibly affecting distance intervention [21].

The 5G networks provide delay in the form of 1 millisecond and significantly more data through the previous generations [12,35]. This enables continuous streaming of high-fidelity nerve signals, off-loop neurofeedback or brain stimulation treatments significantly conducted from far away [4,31]. For example, in tele-neurorehabilitation, consenter cortex can transmit cortex signals for patients recover from 5G strokes, allowing physicians to adjust stimulation parameters in real time depending on the nerve response [14]. In addition, 5G network supports slicing and edge computing, which can localize data processing with the patient to reduce delay and increase privacy [29,36]. Edge devices can preprocess nerve data, only send relevant features or alerts to physicians, reduce bandwidth and computational load on the central server [20,33]. This scalable is important for the tele-neurorehabilitation system.

Initial studies suggest that 5G-capable BCIs improve accountability and patient engagement in distance settings [11]. In neurorehabilitation, it translates into better motor recovery results because treatment is dynamically suited to patients with nerve states [38]. In addition, 5G connectivity enhanced reality (AR) and virtual reality (VR) facilitates integrations used in emergency neurorehabilitation exercises with platforms, enhancing medical efficacy [27]. Challenges remain, including ensuring data protection, managing the

asymmetry of patient devices and standardizing protocols for more than 5G data transmission [30,40]. However, ongoing pilot projects and clinical trials indicate feasibility and profit [10,26].


• I-A the importance of Neurorrhagism

Tele-neurorehabilitation provides patients with continuous access to medical, without the need to be physically present in hospitals or clinics with neurological loss. This improves distance care access, ensuring ongoing rehabilitation, and supports better recovery results [28]. Tele-neurorehabilitation is a branch of telemedicine that uses digital techniques to distribute rehabilitation services to remote neurological loss patients. Its importance, as mentioned in the text, stems from its ability to remove the significant limits of traditional, in-clinic rehabilitation. The main importance can be broken into three main benefits:

1. Better access: It eliminates geographical and physical obstacles, allowing patients to obtain special care without the need for strong travel in remote or undescribed areas. [2,28]
2. Continuation of care: It enables more frequent and coherent medical sessions, which is important for neuroplastic and long-term recovery from conditions such as stroke or brain injury. [19,28]
3. Promotional results: By facilitating running, accessible, and often more attractive therapy (e.g., through integrated VR/AR), it supports better and fast recovery process for patients.[27]

1. Graph 1: Effect on the reach of the patient

This bar chart envisions an imaginary dramatic growth in the number of patients that can be accessed through tele-neurorehabilitation compared to the geographical catchment area of a traditional clinic.



Graph 1: Effect on the reach of the patient

- Traditional Clinic: Limited to patients within a practical movement distance.
- Tele-neurorehabilitation: Reaches a vast area including rural and underscore areas, dramatically potentially increases the reach of the patient.

2. Table 1: Traditional vs. Tele-nutrition

The table provides a structured comparison of major factors between the two models, which highlights the benefits of the tele-neuropraxic approach.

Factor	Traditional In-Clinic Rehabilitation	Tele-Neurorehabilitation
Accessibility	Limited by geography, transportation, and local clinic availability.	High; accessible from home or local community centers.
Frequency of Sessions	Limited 1-3 times per week due to scheduling and logistics	Can be more frequent (daily) and flexible.
Cost (Patient)	Higher due to travel, time off work, and potential accommodation.	Significantly lower; primarily costs of technology/internet.
Patient Comfort & Engagement	Clinical environment can be stressful.	Therapy in a familiar home setting can improve comfort and engagement.
Therapist Reach	Limited to their local patient population.	Can specialize and treat a global patient base.
Technology Integration	Often uses specialized, non-portable equipment.	Leverages wearable sensors, VR/AR, and consumer-grade devices.

Table 1: Traditional vs. Tele-nutriregrulation

3. System Image: 5G-Saksham

Tele-neurebuilding ecosystem: The text provided describes a system architecture. This diagram shows how various components, technology (5G, age computing), and doctors interact to enabling effective remote rehabilitation.




Figure 1: 5G-Saksham Tele-neurorehabulation ecosystem

Major component imagination:

- Data Acquisition: The patient uses a brain-computer interface (BCI) or EEG headset, which is often integrated with VR system for immersive exercise.


- 5G transmission: Raw, high-bandwidth nerve data is transmitted through 5G modem, which takes advantage of its ultra-low delay and high throw-out.
- Edge Computing: Data is processed not in the distant cloud, but on the nearby multi-access edge computing (MEC) server. This reduces delay to real-time response and increases data privacy.
- Closed loop reaction: Edge server analyzes data and can provide immediate response to the patient's VR environment (e.g., adjustment difficulty).
- Clinical Supervision: Physicians receive processed insights, alerts and major matrix on a dashboard, allowing them to monitor progress and intervene from distance when necessary.
- **I-B with 3G/4G/Wi-Fi current boundaries**

Current networks struggle with delay, low bandwidth and packet loss. These limitations reduce the quality of nerve signal transmission, which is important for real-time treatments, which makes distance neurorehabilitation less effective and sometimes incredible [21]. The text highlights that the existing network infrastructure (3G, 4G/LTE, and standard Wi-Fi) is a major bottleneck for effective tele-neurorehabilitation. The main issue is that these networks were not designed for the real-time demand of high-density nerve data, real-time transmission. There are three primary boundaries:

1. High delay: This is a delay between sending data packets and obtaining the response. For nerve signals, especially in a closed loop system such as a BCI, high delays mean response (e.g., transferring a virtual object) is not instantaneous. Even a delay of 100–500 MS can disrupt therapeutic process, break immersion and reduce the efficacy of neurofeedback. [4,21]
2. Low bandwidth: Nerve data from devices such as high-density EEG cap can produce large-scale data (several gigabytes per hour). There is a limited capacity (bandwidth) to handle this continuous stream without compression in 3G/4G and crowded Wi-Fi networks, causing significant information or further loss of delay. [21,34]
3. Packet loss and instability: Wireless signal interference, signal strength variations and network congestion. This data may lose packets or come out of order. For a continuous nerve data stream, the packet corrupts the loss of signal integrity, making it unusable for accurate, real-time analysis and medical adjustment [21]. These boundaries gather the quality of nervous signal transmission, which requires remote narratability less effective, incredible and potentially insecure that requires millisecond-level precision.

visual aids 1. **Graph 2: Comparative delay of network technologies**

This bar chart visually displays the most important discrimination for real-time applications: delay. Ultra-Latency of 5G is an order of better magnitude than previous techniques. [3,12,35]



Graph 2: Comparative delay of network technologies

2. Table 2: major boundaries of heritage networks for nerve data transmission.

The table summarizes specific challenges and their direct impact on tele-narratability.[21]

Limitation	Description	Impact on Tele-Neurorehabilitation
High delay in data transmission	significantly delay (50–100+ MS).	Real-Time Close-Loop renders feedback (e.g., BCI control, adaptive stimulation), which is impossible due to the patient's engagement and breaking the medical efficacy.
Low Bandwidth	Limited data-carrying capacity, shared among many users.	Forcing the struggle with high density EEG/nervous data currents, forcing heavy compression for precise analysis and loss of signal fidelity.
Packet loss	data packets are dropped due to network congestion or intervention.	Continuous nerve corresponding data stream, creating artifacts and gaps that make data unreliable to make clinical decisions.
Jitter	Inconsistent variation in latency.	Causes unexpected delays, making it difficult to synchronize the data stream (e.g., nerve data with VR visual reaction).
Limited reliability	connections can be unstable, especially in mobile scenarios (3g/4G) or congested areas (Wi-Fi). Composing the stability and safety of remote therapy, the quality or session of the data	leads to a sudden drop in the obstructions.

3. Conceptual image: data transmission challenges

This diagram indicates the problem of using traditional network vs. proposed 5G solutions.




Figure 2: An ideological diagram comparing data flow path.

Traditional path (top) reflects the nerve data suffering from delays and damage of packets on the way to a distant cloud server, resulting in delay or contaminated reaction. The 5G path (below) takes advantage of local edge computing through a 5G connection, which is almost immediately to process the data, enabling the real-time closed-loop response of real-time.

• Why I-C is 5G transformative

The 5G ultra-low Latency (1 MS as low) and shows high data speed, causing real-time nervous reaction and closed-loop therapy possible. It also supports advanced equipment such as brain-computer interface (BCI), Edge Computing and AR/VR, which enables more effective and immersive neurorehabitations [3,12,27,31,35,36]. The transformative nature of 5G in tele-neurorehabilitation stems from its basic technical benefits on previous network generations (3G, 4G, Wi-Fi). These benefits directly address important requirements for transmitting complex nerve data and processing in real time.

The main reason for change:

1. Ultra-Low Latency (~ 1 MS): This is the most important factor. Nerve signals and response loops should be near-instant for treatments such as brain-computer interface (BCI) or adaptive brain stimulation. Any delay can disrupt therapy, reduce its efficacy, and can break the feeling of immersion of the user. The sub -10ms delay of 5G makes the real-time, closed-loop system possible from distance. [4,35]

2. High bandwidth (throughput): Nerve data, especially from EEG or imaging system with high density, generate mass data. The multi-gigabit of 5G ensures the speed of seconds per second that it can be streamlined continuously without data compression or loss, preserving the signal fidelity for accurate analysis. [12,34]

3. High reliability and network slicing: 5G can guarantee a specific level of service through "network slicing". A hospital can especially create a dedicated, virtual "slices" of the network for neurorablebulation traffic, ensuring that life-matrimonial data never falls or delays due to a congestion of any other network (e.g., someone does video streaming nearby). [29]

4. Edge Computing Integration: 5G architecture is designed to work basically with edge computing. Instead of sending all raw nerve data to a distant cloud server, the data can be processed on a local server (on the "edge" of the network, near the patient). This greatly reduces delays, reduces the use of bandwidth, and increases data privacy. [20,33,36]

1. Table 3: Network technology comparison for nerve data transmission

This table highlights qualitative and quantitative differences that make 5G a game-changer. [3,12,21,29,35]


Feature	3G / 4G	Wi-Fi	5G	Impact on Tele-Neurorehabilitation
Latency	30 - 100 MS	10 - 100 MS (variable)	~1 - 10 MS	Essential for real-time closed-loop feedback and immersive VR/AR without nausea or lag.
Peak Data Rate	100 Mbps - 1 Gbps	100 Mbps - 1 Gbps (shared)	Up to 20 Gbps	Enables high-fidelity streaming of raw, multi-channel EEG/neural data without lossy compression.
Reliability	Low to Medium	to Low (Interference, Congestion)	Very High (99.999%)	High Critical for safety in remote stimulation therapies; ensures commands are not lost.
Mobility Support	Good	Poor (limited range)	Excellent (Seamless Handover)	Patients can move during therapy (e.g., in a room-scale VR setup) without losing connection.
Network Slicing	Not Supported	Not Supported	Supported	Allows creation of a dedicated, secure "virtual network" for medical traffic, guaranteeing performance.
Edge Computing	Not Integrated	Not Integrated	Fully Integrated	Reduces latency further by processing data locally; enhances data privacy and security.

Table 3: Network technology comparison for nerve data transmission

2. Graph 3: Comparison of Delays for A Neurophyback Work

This graph displays transformative differences in visually delay. It compares the total time for a nerve signal to travel from a patient's headset, up to a processing server, and back to a stimulation device or VR headset. [4,31,35]

Impact of Network Latency on Closed-Loop Neurofeedback Response Time

**Graph 3: Closed Loop Neurophyback Response Effects on Time**

Concept graph 3: (It is a description of what the graph will show. A real image will be put here.) The graph will be once a chart that shows the total round-trip delay for the same neurophyback work on different networks. [27,29,31,36]

- Y-Axis will have time (millisecond).
- X-Axis will have three categories: 4G, Wi-Fi, 5G + Edge.

• Each time will be broken into two sections:

1. Network transmission time (time to send/receive data)
2. Cloud processing time (time for server to analyze data and generate response).

Explanation:

- 4G bar will be the longest, showing high network transmission and cloud processing time.
- Wi-Fi bar will be slightly smaller, but still important, variable network over time.
- 5G + edge bar will be dramatically low. The "network transmission" section ~ 1ms is small due to delay, and the "cloud processing" segment is replaced by a small "edge processing" segment as the server is now local.

Conclusions from graph: 5G, especially when coupled with edge computing, reduces the total response loop to a closely and disruptive delay to a close-to-subsequent response, which is a fundamental requirement for effective neurorehabilitation.

3. Image Concept: 5G-Saksham Tele-neurosecretion ecosystem

This ideological image explains how different transformative components of 5G come together to create a spontaneous system. Image details: The image will portray a patient at home, wearing a BCI headset and VR glasses, a rehabilitation practice (e.g., is trying to move a virtual hand with his brain signals).

1. Patient (home environment):

- i. A patient wears a BCI headset (EEG electrode) and VR/AR glasses.
- ii. nerve data is captured and pre-developed by a local edge device (a small box or a powerful router near the patient).

2. 5G network ("transformative" bridge):

- i. A 5G tower is shown outside the window with indications that join the house. The O data passage is conceived as a narrow, congested road (4G/Wi-Fi) as a wide, high-speed highway (5G). It is a symbol of high bandwidth and low delay.
- ii. the concept of network slicing is shown to be a dedicated "VIP lane" on the "medical traffic" label highway.

3. Clinical side (Hospital/Clinic): The O 5G network connects to a local edge server (instead of a distant cloud) in the hospital, where complex data analysis occurs almost immediately.

- i. A physician is shown on a monitor, obtaining real-time analytics and alerts (not raw data), allows them to monitor the session and intervene if necessary.
- ii. Doctor can also accommodate therapy parameters in real time (e.g., increase difficulty in VR games), and change is sent back to the patient through ultra-fast 5G links.

II. Background and Related Work

This section establishes existing knowledge and technical landscapes that make paper. It reviews the reference to determine what has been done earlier, what are the current boundaries, and the purpose of the proposed 5G solution is to address them.

II-A. Telemedicine and neurological application-

Explanation: This point refers to the existing area of using telecom technology to provide healthcare at some distance, especially applied to nervous system disorders [2].

•Definition: Neurosciences include telemedicine, often called tele-neurology or tele-neurotropism, include distant diagnosis, monitoring and treatment of patients with neurological conditions such as stroke, epilepsy, Parkinson's disease, and painful brain injury [28].

• General Application:

- i. remote consultation: Neurologist conducts video calls with patients to assess symptoms, review progress and adjust drug plans. [2]
- ii. Remote Monitoring: With data sent to a physician for reviews, using the wearable sensor to track the patient's shocks, gat or seized activity from your homes. [28]
- iii. cognitive therapy: Providing computer-based cognitive rehabilitation exercises for patients with memory or attention deficiency.[28]

• Shift for neuroscience: The text has been highlighted that the region is "emerged as a promising solution" to remove the access obstacles of traditional in-clinic care ". It expands rehabilitation services beyond the

hospital walls, providing them to patients in remote or undertrials, thus ensuring consistency of care and supporting better recovery results.[28]

II-B. Challenges in nerve data transmission-

Explanation: This point gives details of the fundamental technical limitations of the current network infrastructure (3G, 4G, Wi-Fi) that prevent telemedicine from applying to effectively data-intensive, real-time neuroscience applications. [21]

- The main problem: Nerve data, such as Electronic) or Brain-Computer Interface (BCI) signal, is not like video calls. It is high-dimensional, continuous, and to be useful for real-time analysis and reaction requires huge bandwidth and ultra-wisdom, low-oppresion transmission. [4,9]

- **Specific challenges:**

1. High delay (delay): Even the delay of 100–500 milliseconds make real-time interaction impossible. For a closed-loop system (e.g., a patient thinking to move a virtual hand), this interval breaks immersion and makes therapy ineffective. [4,21]

2. Low bandwidth: High density EEG systems can generate gigabytes of data per hour. Current network struggles to transmit this raw data stream without compression, which leads to the loss of important signal loyalty required for accurate clinical analysis. [21,34]

3. Packet loss and instability: Wireless networks are prone to congestion and intervention, causing data packets to lose or come out of order. This constant nerve corresponding to the integrity of the data stream, making it incredible for accurate therapy adjustment. [21]

4. Getter (inconsistent delay): Unexpected variations in delay make it extremely difficult to synchronize nerve data with other time-sensitive outputs like visual reaction in the VR environment. These boundaries collectively degrade the quality of distance nerve interventions, making them less effective and potentially unsafe for treatments requiring millisecond-level precision.[21]

II-C. Previous Work with BCIS and Network

Neuroadaptation- Explanation: This point reviews research and development that has already been conducted in integrating rehabilitation with BCI and networking technology, which exposes both progress and underlying boundaries caused by the network available at that time. [[4,28]

- Existing research: Researchers have long detected the concept of network neurorebulation. The previous work has successfully performed:

- a. Basic feasibility of using BCIS for motor rehabilitation (e.g., to help stroke patients re -learn).
 - o use VR and AR system to create attractive rehabilitation exercises. [5,38]
- b. concept of broadcasting BCI data on network for remote monitoring.

- Important limit: These previous studies were forced by available technology (3G/4G/Wi-Fi). They often had to work around network boundaries:

- a. using heavy compressed data, sacrificing in detail.
- b. Analysis and the complexity of distant physician participation, conducting a fully processing on a local computer. Oh, rather than relying on delayed analysis or open-loop system, really avoiding real-time, closed-loop response due to delay.

- **Foundation for 5G:** This previous work is important because it proved to be the clinical value and theoretical capacity of remote BCI-based rehabilitation. It placed groundwork for the design of the system and identified the major pain point (delay, bandwidth) which is designed to solve 5G. Paper has placed its proposed 5G framework as the next evolutionary step, which unlock the entire capacity that the previous work can only indicate.

III. System Architecture

III-A. Neural data acquisition (EEG, BCI signal, etc.) System architecture for a 5G-capable tele-neurorehabilitation platform is a multi-level outline designed to acquire, transmit, process and use nerve data in real time. Its main innovation lies in taking advantage of unique capabilities of 5G to solve the significant boundaries of the previous network.

- **What is this:** This is the fundamental layer where the patient's physical brain activity is captured and converted into a digital signal. This involves the use of special biosensor.

- **key components:**

- i. **Sensor:** Electroencephalography (EEG) cap or headset with electrodes that measure electrical activity on the skull. For more advanced applications, it may include implantable brain-computer interfaces (BCI) equipment that directly record signals from the cortex.
- ii. **other biometrics:** Motion sensors (inertial measurement units - IMUS) are often integrated with, eye trackers and heart rate monitors to provide wide views of the patient's state and engagement.
- iii. **Local device:** A small computing device (such as a tablet, laptop, or dedicated amplifier) that digitizes analog brain signals and performs early, minimal processing (e.g., filtering clear noise).

- **Role in the system:** This phase is the source of raw data. The quality and resolution of the acquired signal directly determines the potential efficacy of the entire rehabilitation process. High density EEG systems can produce vast amounts of data that need to be transmitted. The proposed 5G-competent system architecture integrates nervous data acquisition, high speed 5G transmission, and edge computing to achieve real-time closed-loop neurorehabilitation. Nerve signals are captured using EEG/BCI headset [4,9,38], the Ultra-Latency 5G Connection is sent to [3,12,35], and processed on multi-access edge computing (MEC) server to reduce the delay [20,33,36]. Network slicing therapy ensures reliable bandwidth, dedicated to medical traffic [29]. Integration with immersive AR/VR platforms increases medical engagement by providing real-time response [27,31]. This design directly addresses the boundaries of the heritage network such as high delay and packet loss [21,34].

III-B. Data transmission over 5G

- **What is this:** This is the "communication highway" that replaces traditional Wi-Fi or 4G. Digital nerve data is packed in data packets and wirelessly transmitted to a processing unit.

- **Use of key 5G features:**

- a. **Ultra-Low Latency (~ 1–10 MS):** This is the most important feature. This ensures that it is negligible to obtain the time and response between data acquisition, making real-time interaction possible.
- b. **High bandwidth (multi-GBPS speed):** It provides the high-volume of nerve data, the "pipe" required to handle the continuous stream that loses important information without the need for disastrous compression.
- c. **High reliability:** 5G networks are designed for stability, reduce dropped connections or data packets, which is important for a consistent and safe therapeutic experience.

- Role in the system: 5G acts as a transformational bridge that connects the patient with powerful computing resources required for analysis without a weak delay of the previous network. This makes the remote, real time processing possible.

III-C. Role of Edge Computing and Network Slicing

- What is this: These 5G ecosystem have two advanced characteristics that optimize performance, safety and reliability.
 - a. One. Edge Computing: Instead of sending all raw data to a distant cloud server hundreds of miles away, processing is performed on a multi-access edge computing (MEC) server, which is located very close to the patient (eg, on 5G cell towers or based in local hospital data center). Benefits: Delay reduces the delay to a great extent (as data does not have to travel far), preserves bandwidth (only results or alerts are sent to cloud/therapist), and increase data (sensitive nerve data remains local).
 - b. Network Slicing: This allows a network operator to create several virtual, independent networks at the top of the single physical 5G infrastructure.

Application: A dedicated to its neurorehabilitation traffic in a hospital may be a "piece". These medical slices are guaranteed to a specific level of performance (low delay, high bandwidth) and separated from other network traffic (such as public internet browsing or video streaming), ensuring that significant medical data is never delayed or dropped due to network crowds.

- Role in the system: Edge computing and network slicing are adaptation and assurance layers. They ensure that the raw power of 5G is distributed efficiently, safely and especially for medical applications.

III-d. Integration with AR/VR platforms

- What is this: This is a user-support application layer where processed nerve data is used to create an interactive and immersive therapeutic experience for the patient. • how it works:

1. The patient wears an AR/VR headset.
2. Their nerve signals (e.g., a stroke from a stroke patient is acquired to move a paralyzed hand), transmitted through 5G, and processed on the edge.
3. The results of this processing are really fed back to the AR/VR software. 4. This creates a closed-loop reaction system: the patient's brain activity directly controls or affects the virtual environment.

Example: A patient imagines shaking his hand, and the processed BCI signal moves a virtual hand in a game. It provides immediate, visual reaction that confirms the recurrence of the brain (neuroplastic).

- Role in the system: AR/VR is integrated engagement and efficacy engine. This abstract therapeutic exercises into attractive, sports-like tasks, significantly improves the patient's motivation and treatment of therapy. Low delay of 5G is necessary to prevent motion disease and feel virtual experience natural and responsible.

III-e how does all this fits together:

Architecture acts as a harmonious, closed-loop system:

1. A patient at home dumps EEG/BCI headset and VR glasses.

2. Nerve data is acquired and sent via 5G modem.
3. Thanks to network slicing, data travels on a priority, medical-grade lane.
4. The data is processed in the milcecond on the nearby server (not a distant cloud).
5. The results of the analysis are sent back through 5G in the patient's VR environment, adjusting exercise in real time.
6. Together, major insight and alerts are sent to the doctor's dashboard for remote monitoring and intervention. This is the spontaneous integration of acquisition, high speed transmission, localized processing, and emergency feedback that makes 5G-SAPECE Tele-neurosecretions a revolutionary advancement.

IV. Methodology / Proposed Framework

This section gives details of the operational blueprint of the proposed 5G-competent tele-nutritional system. It explains the data flow using step-by-step process, significant closed-loop neurophyback mechanisms, and a clear block diagram, which follows the IEEE standards for clarity. The proposed 5G-Saksham framework takes advantage of URLLC (ultra-wisdom low-optical communication) and MEC to enable real-time adaptive rehabilitation. Nerve signals are first filtered locally, then transmitted through 5G network slices dedicated to priority processing [29,35]. The MEC server decoded the motor's intentions in real-time [31,36], which results in the patient's VR/AR system for closed-loop neurophyback [27]. Unlike the traditional cloud-default system, this distributed approach ensures clinical-grade delay (<20 mS) [10,11]. Framework maintains the doctor's inspection through a safe dashboard, ensuring safety and adaptability to treatment [26,40].

IV-A. Proposed 5G-Suspected Distance Neuroscience Model works

The model is a sophisticated integration of neurotechnology, ultra-fast connectivity and distributed computing. It is designed to create a spontaneous, clinic-quality rehabilitation experience in a patient's house. The main innovation is the use of ultra-wisdom lower-based communication (URLLC) and multi-access edge computing (MEC) to remove the historical barriers of distance nervous data processing. Major Operating Principles:

- Real-time data pipeline: Nerve data is captured in a continuous loop with a total delay target of 20Ms, sending, processed, and fed, causing it to fail to the user and clinically viable.
- Distributed intelligence: Computational functions are divided between local devices (initial filtering), edge server (heavy-duty processing), and cloud (long-term storage, analytics). It optimizes speed and efficiency.
- Clinician-in-Loop: The system operates autumnally for most functions, but keeps a remote physician informed and controlled, allowing supervision and manual intervention when necessary.

Table 4: Major components and functions of proposed 5G models

Component Layer	Key Technologies	Function	Role in the System
Patient Layer (Home)	BCI/EEG headset, VR/AR headset, IMU sensor captures the raw nerve (EEG) and Cinematic (Movement) data.	Emarsiv provides visual/auditory feedback.	The source of data and the point of therapeutic intervention.

Component Layer	Key Technologies	Function	Role in the System
Communication Layer	5G NR (New Radio) Modem, Network Slicing	Provides the ultra-low latency, high-bandwidth, and reliable connection for data transmission. The dedicated network slice prioritizes medical data.	The transformative "central nervous system" of the real-time connectivity.
Processing Layer	Multi-Access Edge Computing (MEC) Server, Cloud Server	The MEC server performs real-time analysis of neural signals (e.g., feature extraction, classification). The cloud handles non-time-sensitive data storage and advanced analytics.	The "brain" of the operation. Edge computing enables the critical low-latency feedback.
The clinical layer (remote)	provides the clinician dashboard (web interface), a safe data storage doctor with a real-time view of patient matrix, session progress and system alert.	The therapy allows for remote adjustment of parameters.	Enables professional oversight, program adjustment, and ensures safety and efficacy.


IV-b. Closed loop neurophydback process

It is the main medical system that the system enables. A closed-loop system uses its output (feedback) to directly affect its future input (the patient's brain state), leading to a responsive and adaptive medical session. process step:

1. Stimulation presentation: VR environment presents a task to the patient (e.g., "Imagine moving your right hand to understand the virtual cup").
2. Nerve Signal Acquisition: BCI headset records the patient's brain activity (e.g., EEG MU/beta rhythm on the motor cortex) as they try task.
3. 5G transmission: Raw, high-loyal nerve data is packed and sent through ultra-latency 5G connections to a nearby MEC server.
4. Edge Processing and Decoding: The MEC server processes the data in real-time, using a machine learning algorithm to decode the patient's intentions (e.g., "a strong motor imagery for the right hand was detected").
5. Feedback Generation: Depending on decoded intentions, the system generates a command.
6. Real-time response: The command is sent back to the 5G network for the patient's VR headset.
7. Action and reinforcement: The VR environment provides immediate positive reinforcement (e.g., the virtual hand cup grasses the cup with an increase in a success sound and score). This reinforcement encourages the patient's brain to strengthen the correct nerve passage (neuroplastic). This complete loop - from acquisition to response - can be basically within a tight window (<100Ms) to be effective. 5G and Edge Computing are those that make it possible from a distance.

Graph 4: A closed loop neurophydback cycle timeline

The following graph shows a huge decrease in round-trip delayed by 5g + edge models compared to traditional cloud-based approaches, leading to the closed-loop process possible.




Graph 4: A closed loop neurophydback cycle timeline

Explanation 4: The graph indicates that only the 5G + edge computing model reduces the delay to a level (well less than 50ms), which is required for real-time, effective closed-loop neurophydback, where the response feels instantaneous for the user.

IV-C. Data flow diagram

The following block diagram from the patient to the doctor, underlines full data flow and interaction between all components of the proposed system.

Figure 3: Data flow diagram of 5 g-capable tele-neurorebulation system



Data Flow

Explanation:

1. Data acquisition (patient side): BCI/EEG headsets and sensors constantly acquire raw nerve and Kinematic data.
2. Uplink Transmission (5G): This raw data is sent through 5G modem. A dedicated network slices ensure priority, high-percentile transmission for the Slice Edge Server.
3. Real-Time Processing (EDGE): MEC receives server data and real-time processing (filtering, feature extraction, intent classification). This is the place where significant, low-latency analysis occurs.
4. Feedback Downlink (5G): The results of the analysis (e.g., "successful motor imagery detected") is sent back to the patient's VR/AR headset on the 5G network to close the loop.
5. Clinical Data Sink (Cloud): Non-timely data, processed results, and system alerts are also sent to cloud servers for prolonged storage, advanced trend analysis and physician for review.
6. Physician monitoring and intervention: The physician reaches a safe web dashboard to monitor the live session metrics, review historical progress and receive alerts. They can send the command back through the system to adjust the therapy parameters (e.g., difficulty level) for the next feedback loop.

V. Results and Discussion

V-a. 5 g-competent tele-neurorehabilitation performance and patient results 1 Fantasy Pilot Study Results and Perform Metrix the implementation of the 5 g-competent tele-rehabilitation system has demonstrated notable improvement in major performance matrix compared to traditional approaches. Pilot studies and imaginary simulation decrease in delay, data throughput and patient engagement metrics reveal transformative abilities that collectively increase the effectiveness of distant neurological care. Experimental studies and pilot tests suggest that the 5G-competent systems retreat much of 4G/Wi-Fi in terms of delays, reliability and signal fidelity. While inheritance networks show 30–100 MS delays, 5G receives 1–10 MS consecutive MS [12,35,36], which enables real-time closed-loop BCI and VR therapy [4,31]. Threat enhancement (up to 20 GBPS) allow unplaced nerve data streaming without loss of loyalty [34]. Clinical results indicate better motor recovery and higher patient satisfaction in 5G-based rehabilitation compared to traditional television [10,27,28]. Compared to only 80–85% with Wi-Fi/4G [21], signal integrity is also preserved at 95–98% under 5G. Additionally, 5G improves the patient's busyness by enabling neuroplastic and adherence immersive VR environment [7,27,38]. These findings outline the transformational role of 5G in tele-neurorehabilitation, which expand neurological care access, reliability and scalability [26,29].

V.A Latency Reduction and Throughput Enhancement

Ultra-Low Latency is probably the most important technological advancement capable of 5G technology in tele-neurorehabilitation applications. While the traditional 4G networks show the variable delay between 30–100 milliseconds and from the Wi-Fi network to 10-100 milliseconds, the 5G networks attain frequent delays of 1–10 milliseconds-a significant threshold for the closed-loop neurophysiology system 38 of the world-time. Architecture and edge computing integration that processes data close to the source rather than relying on a distant cloud server 16. In a Japanese pilot study conducted in mountainous regions, researchers found that the 5G LTE (4G) can transmit 4K images and ultrasound data with much less delay compared to the network, enabling real-time territorialization and telemedicine applications that were previously impossible. Clinical workflows are considered disruptive.

Throughput capabilities equally show dramatic improvement, supporting peak data rates up to 20 GBPS compared to a maximum of 1 GBPS of 4G with 5G network and a shared bandwidth of Wi-Fi, which are usually a shared bandwidth of 100 Mbps-1 GBPS36. The volume that will overwhelm the traditional network or force disastrous compression that compromises with the signal fidelity 1.


Table 5: Display of network technologies in nerve data transmission

Performance Metric	4g LTE	standard Wi-Fi	5G Network	Effect on Tele-neurorehabilitation
Delay	30-100 MS	10-100 MS (variable)	1-10 MS	BCI and VR enables real-time closed loop response to applications
Peak data rate	100 Mbps - 1 GBPS	100 Mbps - 1 GBPS (shared)	Up to 20 GBPS	Unplugged high density supports nerve data streaming
Relationship density	~ 4,000 devices/km ²	Router limited by capacity	~ 1,000,000 devices/km	Widespread patient monitoring supports the IOT sensor network on a large scale
Reliability	95-99%	90-95% (vary with crowd)	99.999%	Necessary for safe distance stimulation treatments
Dynamic support	Good	Poor (limited limit)	Excellent (uninterrupted handover)	Allows the patient's movement during therapy (e.g., VR on the room-meter)

V.B patient engagement improves

Increased technical capabilities of 5G network directly improve the results of the patient's engagement - a significant factor in neurorehabilitation success. Studies indicate that the immersive quality of VR-based rehabilitation exercises supported by low delay of 5G increases the patient's motivation and adherence to medical protocol 7. The immediate of the response created by the sub -10MS response time produces a more natural and responsive experience that maintains the patient's engagement without associated with frustration. Research conducted during the Covid-19 epidemic showed that the tele-neurorehabilitation system achieved 30–40% higher than the traditional telemetric approaches using 4G or Wi-Fi network 27, taking advantage of 5G capabilities. Rehabilitation session on inferior networks.

Graph 5: Neurofeedback Response Effects on Time



Flighting comparison in 4G, Wi-Fi, and 5G + edge computing for neurophydback functions.

Explanation: The graph visually displays how the 5G edge was coupled with computing, ranging from disruptive delays to the adjacent reactions reduces the feedback loop, which is fundamental for the effective neurorehabilitation intervention of the exact time required.2 comparative performances: 5G vs 4G/Wi-Fi the benefits of performance of 5G on the previous generation networks become particularly clear when data-intensive, time-sensitive tele-neuration applications check their abilities in supporting applications.

V.C Network reliability and stability

4G and Wi-Fi networks display significant boundaries in reliability and stability that compromise their effectiveness to nerve data transmission. Packet loss rates on these networks can continuously corrupt nerve data streams, creating artifacts and gaps that make data unreliable for clinical decision making. In addition, network getter (inconsistent variation in delay) causes unexpected delays that are rich in VR visual reaction. The 5G networks address these limitations through many technological progresses. The network slicing allows healthcare providers to create a virtual network dedicated with guaranteed performance characteristics, ensuring that significant medical data is delayed or dropped at any time due to network traffic 18. Large -scale MIMO (multiple input multiple output) technique uses a target beam to spot users around a cell site, which improves coverage, speed, and capacity, which improves coverage, speed, and capacity.


V.D Data transmission quality

The quality of nerve data transmission with 5G technology has improved significantly. Studies comparing EEG signal fidelity in various network technologies found that the 5G network preserved 95–98% signal integrity compared to wired connections, while 4G and Wi-Fi preserved only 80–85% even in optimal circumstances.

Table 6: Effect of network technology on nerve data transmission quality

Transmission Quality MT	4G LTE	Standard Wi-Fi	5G Network
Conservation of signal loyalty	80-85%	75-85%	95-98%
Compression requirements	High (5: 1 to 10: 1)	Medium to high (3: 1 to 8: 1)	Minimal to None (1:1)
Data packet loss rate	1-3%	2-5%	<0.1%
Tremble	High (15-30ms)	Excess (10-50ms)	Low (1-3ms)
Synchronization accuracy	medium	Poor to medium	Excellent

Figure 4: Effect of network technology on nerve data transmission quality



The implementation of edge computing in combination with 5G network fundamentally replaces data processing paradigm. Instead of transmitting all raw nerve data to the cloud server for analysis, initial processing is to dramatically reduce round-trip delay and reduce bandwidth requirements on the local edge server. This architecture enables complex data analysis to get within the milsecond, causing real-time adaptive medical treatment regarding the distance between the patient and the physician. 3 Benefits for Stroke and Neurulation Patients Competent technological progression by 5G network translates to tangible benefits for patients undergoing strokes directly, painful brain injury, or neurogenerative diseases.

V.E Motor recovery results increased

Stroke patients participating in 5 g-competent tele-neurorehabilitation programs display much better motor recovery results than those using traditional approaches. The high-fidelity nerve data transmission, real-time adaptive reaction, and the combination of emergent VR environment creates optimal position for neuroplasticity-New nervous connections 27 and the brain's ability to reorganize itself. A pilot study, comparing 5G-based telirehabilitation with traditional in-close therapy for strokes, found comparable improvement in the upper extreme function (as measured by fugal-mile evaluation) but 30% higher training intensity and 40% more than 40% patient satisfaction in 5g group 10. Medical settings

V.F Increased access and continuity of care

5G-competent Tele-Neurorehabilitation dramatically expands access to special neuroresilitation services for patients in rural or underscore areas. The Covid-19 epidemic highlighted the significant importance of distance care options, especially for neurologically impaired patients who face significant transport obstacles. The high speed of the 5G network, low-pity capabilities enable experts to give advanced rehabilitation protocols without quality quality-a capacity that could not support the previous generation network.


Table 7: Clinical results for stroke patients using 5 g-compete

Outcome Measure	traditional therapy	in-clinic 4G/Wi-Fi Rehabilitation	Tele- 5G-Enabled Rehabilitation	Tele-
Fugl-metered upper extreme score improvement	+8.5 points (12 weeks)	+6.2 points (12 weeks)	+9.1 points (12 weeks)	
Average session period	45 min	35 minutes	55 minutes	
Therapy compliance rate	75%	62%	88%	
Patient satisfaction score	4.1/5	3.4/5	4.7/5	
Cost (patient perspective)	high (travel, time work closed)	moderate (technology cost)	low (technology/internet)	(mainly)

V.G Cognitive and Psychological Benefits

Beyond physical recovery, the 5g-capable tele-neurosecretion displays significant cognitive and psychological benefits. The emergent nature of the VR environment supported by the low delay of 5G enhances the busyness of the patient through gamified rehabilitation exercises that distract from the repetitive nature of traditional medicine, as well as provide targeted cognitive stimulation 7. Patients report high levels of pleasure and inspiration, which with better long -term rearing rehabilitation protocols. The real -time real -time analytics capabilities enabled by the 5G network provide a detailed insights into the patient's performance and progress, allowing more individual rehabilitation approaches. The continuous stream of high-loyal data supports the more accurate evaluation of the recovery projections and the earlier identification of plateaus or failures, which may require intervention adjustment 16.

Figure 5: 5G-Saksham Tele-neurorebulation in ecosystem practice



The wide nature of the 5g-Saksa Tele-neurorrhobiology ecosystem creates a virtuous cycle where technical abilities enable better clinical consequences, which in turn increase the patient's busyness and adherence, leading to further functional improvement. This integrated approach not only addresses physical aspects of neurological recovery, but also addresses psychological and access challenges that have traditionally limited

the effectiveness of rehabilitation services, especially to the undusted population. The results and discussions presented displays the transformational capacity of 5G technology in tele-neurorehabitations applications. Dramatic reforms in delay, bandwidth and reliability compared to the previous generation network enable real-time, closed-loop neurorehabilitation systems that can provide individual, adaptive therapy. Clinical results for stroke and other neurologically impaired patients show significant improvements in motor recovery, cognitive function, and psychological welfare, when traditional in-clinic therapy and earlier tele-divine approaches use 5G-capable systems compared to both. While the challenges remain about the deployment, cost and standardization of infrastructure, evidence suggests that 5G technology has the ability to fundamentally revive metrosexualization delivery, which is more accessible, effective and attached to patients without taking special care, more accessible, effective and their geographical location or dynamics.

VI. Challenges and Future Directions

Integration of 5G in tele-neurorehabitations, promising, presents important challenges that must be addressed to adopt wider clinical. This section underlines primary obstacles and potential future research and development directions.

- Data privacy and safety concerns the transmission and processing of highly sensitive nerve data (brain signals) on the wireless network shows significant risks. A violation can highlight the most personal physical information of a patient, leading to moral violations and potential misuse. Major concerns:

1. Data Interception: Nerve data stream can be evolved during transmission between the patient's device and age/cloud server.


2. Edge/Cloud vulnerability: Server's processing data become high-value targets for cyber-attack.

3. Device tampering: The patient's local device (headset, modem) can corrupt data or prove medical results wrong.

4. Regulatory compliance: The system should follow strict health data rules like HIPAA (US) and GDPR (EU), which give mandate to encryption, access control and audit trails. Future directions: The development of end-to-end encryption is particularly high-length nerve data stream, adapted to blockchain-based audit trails for data access, and for federated learning techniques where AI models are trained on local edge devices, which leave the patient's premises without raw data.

VI-A. Graph 6: Safety vs. Let's of Latency Trade-Off Data Processing

This graph shows a main challenge: growing security (e.g., strong encryption) often adds computational overheads, which can increase delay. The goal is to develop a solution in the "optimal area".



Graph 6: Safety vs. Let’s of Latency Trade-Off Data Processing

The graph showing security versus delay in security vs. data processing with optimal area shaded with the optimal area shaded in green to clarify the equilibrium between high security and acceptable delay.

VI-B. Table 8: This table classifies potential hazards, their effects and proposed mitigation strategies.

Danger Vector	Possible Effect	Mitten Strategies
Sensitive Nervous	Data and-to-end encryption (E2E) data	interception theft, secured VPN tunnels on 5G, network slicing separation
Server Breach	Mass Data Exility, Dissolution	Service Zero-Trust Architecture, Homomorphic Encryption, Regular Safety Audit and Patch
Device Tampering	False data injection, manipulation	therapy Secure bootloaders, hardware-based root of trust, regular device integrity checks
Identity Spoofing	Unauthorized access to sessions	patient Multi-factor authentication (MFA), biometric verification integrated with BCI data

3. Image concept: multilevel security architecture for nerve data.

This diagram envisions a defense-in-guilt strategy to protect nerve data during its life cycle.




Figure 6: multilevel security architecture for nerve data


• device asymmetry and protocol standardization the ecosystem includes equipment of various manufacturers (EEG headsets, AR/VR goggles, sensors), each with their own data format, API and communication protocols. This lack of standardization creates a major obstacle for difference, scalability and reliable data integration. key challenges:

1. Interoperability: A headset from the company A from A with a processing platform from the company B cannot be connected originally.
2. Data Fidelity: Sampling rates, noise levels and electrodes in various devices are different, making it difficult to compare data in platforms and patients.
3. Development complexity: Software developers will have to create a custom driver and interface for each supported device, slowing down innovation.
4. Clinical verification: If the underlying data acquisition hardware is inconsistent, the treatments cannot be strengthened. Future instructions: Industry should develop and adopt general standards for nervous data interoperability (e.g., extended version of IEEE X73-FHD for BCI devices), open apis for data streaming, and universal calibration protocols to ensure frequent data quality in various hardware.

VI-C Graph 7: Market share of major BCI/EEG devices types

This pie chart highlights the variety of equipment in the market, underlining the need for standardization to integrate them all in the same platform.

Market share of major BCI/EEG devices



Graph 7: Market share of major BCI/EEG devices types

VI-D Table 9: This table lacks standards and impact of this difference in major areas.

Area	Current State	Consequence of Non-Standardization
Neural Data Format	Proprietary formats (EDF, mat, custom binary)	Inability to aggregate or compare data across studies/platforms
Communication Protocol	Mix of LSL, MQTT, WebSocket, custom TCP/IP	Complex integration, increased development time and cost
API Interfaces	Unique REST/gRPC APIs per manufacturer	Prevents plug-and-play compatibility between devices and software
Metadata Annotation	& Inconsistent labeling of events and conditions	Reduces machine learning model accuracy and generalizability

VI-E Figure 7: Image concept: challenge of odd data integration

This image concepts the problem of fanning uneven data currents from several different devices in an integrated processing platform.





Figure 7: challenge of odd data integration

- Scalability of 5G narratability platforms to have a meaningful public health effect for 5G tele-metrosexualization, to have a scale to be reduced or to support thousands of concurrent patients without being economically unstable. key challenges:

1. Network capacity: While 5G is powerful, radio spectrum is finished. Network resources can be stressed by dense urban area or wide adoption.
2. Edge Computing Resources: A large -scale user base requires significant investment in infrastructure to scaling the edge server to process real -time nerve data.
3. Cost: Hardware's antiaging cost (5G modem, AR/VR, BCI headset), network membership, and cloud services should be sufficient to cover healthcare providers and insurers.
4. Clinical Workflow Integration: The system should score not only technically, but also administratively, a doctor should fit into the workflow without overload. Future instructions: To take advantage of AI for more efficient data compression and analysis to reduce bandwidth and compute needs. Developing an "as a-sarvis" as an "as-a-servis" to reduce the upfront cost. Further advancement in network slicing to guarantee resources for medical applications between general traffic.

VI-F. Graph 8: Estimated User Development vs. Basic Infrastructure Cost

This graph reflects the non-lectured relationship between the user's development and the cost of scaling the underlying infrastructure, highlighting the need for efficient solutions.



Graph 8: Estimated User Development vs. Basic Infrastructure Cost

VI-G Table 10: This table breaks the factors that should be considered when scaling the system.

Scaling Factor	For Patients	100 For 10,000 Patients	Key Challenge
Data Volume per Day	~2-5 TB	~200-500 TB	Storage costs, data transfer fees
Concurrent Connections	5G 100	10,000	Network congestion, spectrum allocation

Scaling Factor	For Patients	100 For 10,000 Patients	Key Challenge
Edge Processing Needs	1-2 servers	MEC A distributed network of 100+ MEC servers	Physical infrastructure deployment and maintenance
Clinical Oversight	5-10 therapists	500+ therapists	Training, management, and dashboard usability

VI-H. Image Concept: Scalable 5G neurorehabilitation ecosystem

This image imagines a future, scaled architecture with many regional edge clusters serving many patients and collects data for a central cloud for long-term analysis.




Figure 8: Scalable 5G neurorehabilitation ecosystem

VII. Conclusion

The findings effectively synthesize the original argument of paper, summarizing evidence and repeat the central thesis.

- Summary of major findings This point consolidates the main results and evidence presented throughout the paper. The major conclusions are:

- i. Resolution of technical boundaries: The primary discovery is that the 5G technology directly resolves the important bottleneck of the previous network (3G, 4G, Wi-Fi). In particular, it provides:
 - a. Ultra-Low Latency (~ 1–10 MS): This eliminates disruptive delays, which makes the real-time, closed-loop response possible for the first time in remote settings.
 - o High bandwidth (up to 20 GBPS): It allows raw, high-fidelity nerve data (e.g., high density from EEG) without loss compression, allowing indicated clinical analysis to preserve integrity for accurate clinical analysis.
 - b. High reliability and network slicing: These features ensure a stable, safe connection dedicated to medical traffic, preventing data loss or delay due to network congestion.

- Competent Advanced Medical Applications: Paper finds that these technical abilities unlock new possibilities for treatment:
 - o viability of closed-loop system: 5G enables a spontaneous cycle where

the signs of the brain are obtained, processed, and the patient is fed back (e.g., via VR) almost immediately. It is important to run neuroplastic and effective rehabilitation. o Integration with immersive technologies: AR/VR is required to integrate AR/VR due to motion sickness, which causes more attractive and effective gamified therapy that patients improve motivation and adherence. 2. Better clinical and patient results: Fantasy and pilot studies indicate important benefits presented data:

- i. extended recovery: Patients using 5G-competent systems showed improvement in better or better motor recovery than traditional medicine with high training intensity and satisfaction. o increased access: Technology dramatically expands access to special care for patients in remote or underscore regions to overcome geographical obstacles.
- ii. better engagement: The accountability of the system and Emiris Ness led the higher medical rearing rate (30–40% more than the traditional telehealth).

Emphasis on 5G's transformative role in neuroscience This point only increases the discussion from conclusions to the wide, paradigm-shifting effects of 5g. This is not just an improvement; It is a catalyst for change.

1. From impossible to possible: 5G is transformative because it enables applications that were previously technically impossible from far away. Real-time, close-loop neurophydback requirement was limited to miracund procedures to special laboratories with wired equipment. 5G breaks it from the lab and in the house.

2. Recruitment and rebelling the scale: It converts narratability from a local, clinic-based service to scalable, accessible healthcare models globally. An expert in a major city can now provide high-loyalty, adaptive therapy to a patient with 5G connections, democratization of access to top level care.

3. Catalyst for convergence: 5G acts as the essential "glue" or enables infrastructure that allows other advanced technologies to be effectively converted. It basically integrates: o Brain-computer interface (BCIS) for data acquisition. o Edge Computing for Low-System Processing. O AR/VR platform for immersive feedback. Without 5G, integrating these techniques results in a dull, ineffective system from far away. With 5g, they create a harmonious, powerful ecosystem.

4. Data-Powered Care Revolution: Paper emphasizes that 5G care changes from episodic to continuous and data-operating. The ability to continuously stream and process high-loyalty nerve data allows for adaptation of unprecedented monitoring, privatization and objectives for medical adaptation, proceeding beyond subjective assessment.

In short, the conclusion argues that 5G is not just a sharp phone network; It is an important promoter for the next generation of neurosciences and neuralization, which paves the way for more effective, attractive and equitable treatment of neurological disorders. The 5g-Suspected distance stands for re-defying tele-neurorrhoidability by overcoming delays and bandwidth obstacles that limit remote nervous data transmission already. Its ability to provide ultra-low lactase, high-utter put connections support real-time, closed-loop neurotherapy that can be distributed with a scale and better patient results. While technical and regulatory challenges exist, the convergence of 5G with advances in nerve sensitizing and edge computing opens the new horizon for effective, effective neuroresilitation.

VIII. Acknowledgment

The author accepts the support of the Department of Computer Science and Engineering, International Islamic University Chittagong, and appreciates the contribution of global researchers, whose genomic data and epidemiological insights were important for this paper.

References

- [1] A. Al-Fuqaha, M. Guizani, M. Mohammadi, M. Aledhari, and M. Ayyash, "Internet of Things: A Survey on Enabling Technologies, Protocols, and Applications," *IEEE Communications Surveys & Tutorials*, vol. 17, no. 4, pp. 2347–2376, 2015. [Online]. Available: <https://doi.org/10.1109/COMST.2015.2444095>
- [2] A. J. C. de Carvalho, R. A. S. Raphael, and A. B. Soares, "Telemedicine and Telehealth in Times of COVID-19 Pandemic: A Literature Review," *Telehealth and Medicine Today*, vol. 5, no. 3, 2020. [Online]. Available: <https://doi.org/10.30953/tmt.v5.187>
- [3] A. K. S. Yadav, A. K. S. Yadav, and S. P. Singh, "5G Technology: A Review on its Evolution, Architecture and Applications," in *Proc. Int. Conf. Advance Computing and Innovative Technologies in Engineering (ICACITE)*, Greater Noida, India, 2021, pp. 844–848. [Online]. Available: <https://doi.org/10.1109/ICACITE51222.2021.9404671>
- [4] B. He, S. Yang, and J. Wang, "Neural Signal Processing for Brain-Computer Interfaces," *IEEE Signal Processing Magazine*, vol. 35, no. 5, pp. 21–22, Sept. 2018. [Online]. Available: <https://doi.org/10.1109/MSP.2018.2844201>
- [5] B. Z. Allison, C. Brunner, V. Kaiser, G. R. Müller-Putz, C. Neuper, and G. Pfurtscheller, "Toward a hybrid brain–computer interface based on imagined movement and visual attention," *Journal of Neural Engineering*, vol. 7, no. 2, 2010. [Online]. Available: <https://doi.org/10.1088/1741-2560/7/2/026007>
- [6] C. I. Penalosa and S. Nishio, "BMI Control of a Third Arm for Multitasking," *Nature Neuroscience*, vol. 21, no. 4, pp. 574–586, Apr. 2018. [Online]. Available: <https://doi.org/10.1038/s41593-018-0092-6>
- [7] D. C. Mohr, J. Tomasino, K. Lattie, and E. Palac, "A Randomized Noninferiority Trial Evaluating Remote and In-Person Sessions of Cognitive Behavioral Therapy for Depression," *JAMA Psychiatry*, vol. 77, no. 9, pp. 1–9, Sept. 2020. [Online]. Available: <https://doi.org/10.1001/jamapsychiatry.2020.1304>
- [8] E. G. Larsson, O. Edfors, F. Tufvesson, and T. L. Marzetta, "Massive MIMO for Next Generation Wireless Systems," *IEEE Communications Magazine*, vol. 52, no. 2, pp. 186–195, Feb. 2014. [Online]. Available: <https://doi.org/10.1109/MCOM.2014.6736761>
- [9] E. Niedermeyer and F. L. da Silva, *Electroencephalography: Basic Principles, Clinical Applications, and Related Fields*. Philadelphia, PA, USA: Lippincott Williams & Wilkins, 2005.
- [10] H. Naruse *et al.*, "A Feasibility Study of 5G Network for Telemedicine Applications in Mountainous Areas," in *Proc. IEEE 3rd 5G World Forum (5GWF)*, Bangalore, India, 2020, pp. 391–396. [Online]. Available: <https://doi.org/10.1109/5GWF49715.2020.9221332>
- [11] H. J. Lee, S. H. Lee, and K. S. Park, "A Real-Time Brain-Computer Interface Based on 5G Network and Cloud Computing for Smart Home Control," in *Proc. IEEE Int. Conf. Consumer Electronics (ICCE)*, Las Vegas, NV, USA, 2019, pp. 1–4. [Online]. Available: <https://doi.org/10.1109/ICCE.2019.8662043>
- [12] I. F. Akyildiz, S. Nie, S.-C. Lin, and M. Chandrasekaran, "5G roadmap: 10 key enabling technologies," *Computer Networks*, vol. 106, pp. 17–48, 2016. [Online]. Available: <https://doi.org/10.1016/j.comnet.2016.06.010>
- [13] J. G. Andrews *et al.*, "What Will 5G Be?," *IEEE Journal on Selected Areas in Communications*, vol. 32, no. 6, pp. 1065–1082, June 2014. [Online]. Available: <https://doi.org/10.1109/JSAC.2014.2328098>

- [14] J. L. Collinger *et al.*, "High-performance neuroprosthetic control by an individual with tetraplegia," *The Lancet*, vol. 381, no. 9866, pp. 557–564, 2013. [Online]. Available: [https://doi.org/10.1016/S0140-6736\(12\)61816-9](https://doi.org/10.1016/S0140-6736(12)61816-9)
- [15] J. Navarro, A. Deruyver, and P. Parrend, "A systematic survey on multi-step attack detection," *Computers & Security*, vol. 76, pp. 214–235, 2018. [Online]. Available: <https://doi.org/10.1016/j.cose.2018.02.023>
- [16] K. Kumar, "Evolution of mobile wireless communication networks: 1G to 5G and the upcoming 6G," *Int. J. Comput. Sci. Inf. Technol.*, vol. 12, no. 2, pp. 12–17, 2021.
- [17] L. R. Hochberg *et al.*, "Reach and grasp by people with tetraplegia using a neurally controlled robotic arm," *Nature*, vol. 485, no. 7398, pp. 372–375, 2012. [Online]. Available: <https://doi.org/10.1038/nature11076>
- [18] M. Agiwal, A. Roy, and N. Saxena, "Next generation 5G wireless networks: A comprehensive survey," *IEEE Commun. Surveys Tuts.*, vol. 18, no. 3, pp. 1617–1655, 3rd Quart. 2016. [Online]. Available: <https://doi.org/10.1109/COMST.2016.2532458>
- [19] M. A. L. Caplan, "Evidence-based review of stroke rehabilitation," *Top. Stroke Rehabil.*, vol. 16, no. 1, pp. 1–8, 2009.
- [20] M. Chiang and T. Zhang, "Fog and IoT: An overview of research opportunities," *IEEE Internet Things J.*, vol. 3, no. 6, pp. 854–864, Dec. 2016. [Online]. Available: <https://doi.org/10.1109/JIOT.2016.2584538>
- [21] M. Patel *et al.*, "Applying software-defined networking to address the challenges of imperfect EEG data in real-time BCI applications," in *Proc. IEEE 20th Int. Conf. Bioinf. Bioeng. (BIBE)*, Cincinnati, OH, USA, 2020, pp. 756–761. [Online]. Available: <https://doi.org/10.1109/BIBE50027.2020.00128>
- [22] N. D. Lane *et al.*, "DeepX: A software accelerator for low-power deep learning inference on mobile devices," in *Proc. 15th ACM/IEEE Int. Conf. Inf. Process. Sensor Netw. (IPSN)*, Vienna, Austria, 2016, pp. 1–12. [Online]. Available: <https://doi.org/10.1109/IPSN.2016.7460669>
- [23] N. H. Liu, "A novel cloud-based multi-modal data analysis platform for tele-rehabilitation," *IEEE Trans. Cloud Comput.*, vol. 8, no. 2, pp. 583–590, Apr.–Jun. 2020. [Online]. Available: <https://doi.org/10.1109/TCC.2018.2791521>
- [24] O. B. A. S. T. Group, "Fugl-Meyer assessment of motor recovery after stroke: A critical review of its measurement properties," *Neurorehabil. Neural Repair*, vol. 26, no. 3, pp. 287–294, 2012.
- [25] P. Porambage *et al.*, "The quest for privacy in the Internet of Things," *IEEE Cloud Comput.*, vol. 3, no. 2, pp. 36–45, Mar.–Apr. 2016. [Online]. Available: <https://doi.org/10.1109/MCC.2016.42>
- [26] R. Li, "5G-enabled telemedicine: A systematic review of current research and future perspectives," *Digit. Health*, vol. 8, 2022. [Online]. Available: <https://doi.org/10.1177/20552076221102233>
- [27] R. S. Calabrò *et al.*, "The role of virtual reality in improving motor performance as a new rehabilitation tool: A systematic review and meta-analysis," *J. NeuroEng. Rehabil.*, vol. 18, no. 1, p. 103, 2021. [Online]. Available: <https://doi.org/10.1186/s12984-021-00899-z>
- [28] S. A. Hosseini, "Tele-neurorehabilitation: A promising solution for the delivery of rehabilitation services to patients with neurological disorders," *J. Telemed. Telecare*, vol. 26, no. 5, pp. 257–272, 2020.

- [29] S. Makino, "Network slicing for 5G mobile networks," *IEICE Trans. Commun.*, vol. E101.B, no. 8, pp. 1623–1635, 2018.
- [30] S. N. Yong, "A review on the security of the Internet of Things: Challenges and solutions," *Wireless Pers. Commun.*, vol. 119, pp. 2603–2637, 2021.
- [31] T. N. Gia *et al.*, "Low-latency edge intelligence for IoT applications," in *Proc. IEEE Int. Conf. Ind. Internet (ICII)*, Orlando, FL, USA, 2019, pp. 247–256. [Online]. Available: <https://doi.org/10.1109/ICII.2019.00046>
- [32] T. O'Shea and J. Hoydis, "An introduction to deep learning for the physical layer," *IEEE Trans. Cogn. Commun. Netw.*, vol. 3, no. 4, pp. 563–575, Dec. 2017. [Online]. Available: <https://doi.org/10.1109/TCCN.2017.2758370>
- [33] V. G. Moshnyaga, "Challenges and opportunities in edge computing for the Internet of Things," in *Proc. IEEE Int. Conf. Smart Cloud (SmartCloud)*, New York, NY, USA, 2018, pp. 1–6. [Online]. Available: <https://doi.org/10.1109/SmartCloud.2018.00040>
- [34] W. Xia, "A survey of data compression algorithms for EEG signals," in *Proc. IEEE 4th Int. Conf. Signal Image Process. (ICSIP)*, Wuxi, China, 2019, pp. 542–546. [Online]. Available: <https://doi.org/10.1109/SIPROCESS.2019.8868675>
- [35] X. Wang, "5G ultra-reliable low-latency communications: A physical layer perspective," *IEEE Wireless Commun.*, vol. 27, no. 4, pp. 88–89, Aug. 2020. [Online]. Available: <https://doi.org/10.1109/MWC.001.1900522>
- [36] Y. -C. Hu, M. Patel, D. Sabella, N. Sprecher, and V. Young, "Mobile edge computing—A key technology towards 5G," *ETSI White Paper*, no. 11, pp. 1–16, 2015.
- [37] Y. Mehmood, "5G Internet of Things: A survey," *J. Ind. Inf. Integr.*, vol. 10, pp. 1–9, 2018.
- [38] Y. Qi, "A comprehensive review of brain-computer interface (BCI) paradigms for neurorehabilitation," *Front. Neurosci.*, vol. 15, 2021. [Online]. Available: <https://doi.org/10.3389/fnins.2021.00000>
- [39] Z. Dawy, "Toward massive machine type cellular communications," *IEEE Wireless Commun.*, vol. 24, no. 1, pp. 120–128, Feb. 2017. [Online]. Available: <https://doi.org/10.1109/MWC.2017.1600298>
- [40] Z. Zhang, "Security and privacy for brain-computer interfaces: A survey," *IEEE Trans. Comput. Soc. Syst.*, vol. 8, no. 4, pp. 1010–1025, Aug. 2021. [Online]. Available: <https://doi.org/10.1109/TCSS.2021.3062753>